

**Before opening induction  
chambers, first purge with a  
1.5 L/min oxygen flush for >2  
seconds.**

SC #279

EFF 11/22

4" x 6" horizontal template

Inner box dimensions 3.7" x 5.7"

Outer box dimensions 4" x 6"